


Sample Information

Name : Argireline  
 Sequence : Ac-EEMQRR-NH2 (Acetate Salt)  
 Pump A : 0.1% Trifluoroacetic in 100% water  
 Pump B : 0.1% Trifluoroacetic in 100% acetonitrile  
 Total Flow : 1ml/min  
 Wavelength : 214nm  
 Analytical column type : Phenomenex Luna® 5 µm C18 (2) (4.6\*250mm\*5µm)  
 Dissolution method : 100% H2O  
 Inj. Volume : 15ul

Time	Module	Action	Value
0.01	Pumps	B.Conc	10
25.00	Pumps	B.Conc	35
33.00	Pumps	B.Conc	100
38.00	Pumps	B.Conc	100
40.00	Pumps	B.Conc	10
50.00	Controller	Stop	

Chromatogram



Peak Table

Detector A 214nm

Peak#	Ret. Time	Area	Height	Area%
1	5.219	70647	10804	3.366
2	12.833	17490	5299	0.833
3	12.962	2003322	278813	95.458
4	13.140	7179	3176	0.342
Total		2098638	298092	100.000

Mail : [marketing@motifbiotech.com](mailto:marketing@motifbiotech.com)

Web : <https://www.motifbiotech.com>

Add : B04, Building 17, No. 122, Yong'an Rd, Suzhou National Hi-Tech District, PRC