


Sample Information

Name :Dipeptide-6
 Sequence :Pro-Hyp
 Pump A :0.1%Trifluoroacetic in 100% water
 Pump B :0.1%Trifluoroacetic in 100% acetonitrile
 Total Flow :1ml/min
 Wavelength :214nm
 Analytical column type :Phenomenex Luna® 5 µm C18(2) (4.6*250mm*5 µm)
 Dissolution method :100%H2O
 Inj. Volume : 25ul

Time	Module	Action	Value
0.01	Pumps	B.Conc	0.1
20.00	Pumps	B.Conc	20
33.00	Pumps	B.Conc	100
38.00	Pumps	B.Conc	100
40.00	Pumps	B.Conc	0.1
50.00	Controller	Stop	

Chromatogram



Peak Table

Detector A 214nm

Peak#	Ret. Time	Area	Height	Area%
1	7.577	5785	693	0.043
2	8.563	76943	3483	0.577
3	8.748	13124034	1321567	98.355
4	9.046	136807	10196	1.025
Total		13343569	1335939	100.000

Mail : marketing@motifbiotech.com

Web : https://www.motifbiotech.com

Add : B04, Building 17, No. 122, Yongan Rd, Suzhou National Hi-Tech District, PRC