


Sample Information

Name :Palmitoyl Tripeptide-1
 Sequence :Pal-GHK (HCl Salt)
 Pump A :0.1% Trifluoroacetic in 100% water
 Pump B :0.1% Trifluoroacetic in 100% acetonitrile
 Total Flow :1ml/min
 Wavelength :214nm
 Analytical column type :YMC-Triart C18(4.6*250mm*5um)
 Dissolution method :50% ACN+50% H2O
 Inj. Volume : 25ul

Time	Module	Action	Value
0.01	Pumps	B.Conc	40
25.00	Pumps	B.Conc	65
33.00	Pumps	B.Conc	100
38.00	Pumps	B.Conc	100
40.00	Pumps	B.Conc	402
50.00	Controller	Stop	

Chromatogram



Peak Table

Detector A 214nm

Peak#	Ret. Time	Area	Height	Area%
1	12.185	1711	123	0.051
2	12.721	20479	2225	0.606
3	12.867	3314236	509690	98.141
4	13.162	39002	3813	1.155
5	15.572	1583	233	0.047
Total		3377011	516084	100.000

Mail : marketing@motifbiotech.com

Web : https://www.motifbiotech.com

Add : B04, Building 17, No. 122, Yongan Rd, Suzhou National Hi-Tech District, PRC