


Sample Information

Name :Hexapeptide-9
 Sequence :GPQGPQ
 Pump A :0.1%Trifluoroacetic in 100% water
 Pump B :0.1%Trifluoroacetic in 100% acetonitrile
 Total Flow :1ml/min
 Wavelength :214nm
 Analytical column type :Phenomenex Luna® 5 µm C18(2) (4.6*250mm*5 µm)
 Dissolution method :100%H2O
 Inj. Volume : 25ul

Time	Module	Action	Value
0.01	Pumps	B.Conc	5
25.00	Pumps	B.Conc	30
33.00	Pumps	B.Conc	100
38.00	Pumps	B.Conc	100
40.00	Pumps	B.Conc	5
50.00	Controller	Stop	

Chromatogram



Peak Table

Detector A 214nm

Peak#	Ret. Time	Area	Height	Area%
1	8.216	73059	2584	0.344
2	9.442	21059919	1613735	99.292
3	10.277	77037	6612	0.363
Total		21210015	1622930	100.000

Email: marketing@motifbiotech.com

Web: http://www.motifbiotech.com

Add : B04, Building 17, No. 122, Yongan Rd, Suzhou National Hi-Tech District, PRC